



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

803

DATE: Mar 16, 2006 INITIALS: SEC OWNER: WSP
DISTRICT: 8 ADDRESS: _____
LOCATION: 7 M Shop COUNTY: _____
MAKE: Huani TYPE: port wheel weights ACCURACY CLASS: III
MODEL: 101 ID. NO: 21830 CAPACITY: 20,000 lbs SECTIONS: 1
MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: dial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Center</u>	<u>5000</u>	<u>5000</u>	<u>0</u>				
	<u>10000</u>	<u>10000</u>	<u>0</u>				
	<u>15000</u>	<u>15000</u>	<u>0</u>				
<u>Center</u>	<u>20000</u>	<u>20000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by:

David Gorman

Date:

3/28/06

501

State of Washington

County of Thurston

Signed and sworn to (or affirmed) before
me on March 28, 2006 by

Michael J. Evans
(Signature of notary public)

Michael J. EVANS

Printed Name of Notary

My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS